


Cu 327.396 {103} (Axial)

Label	T
BLANK	0.012
Calibration Properties	
0.001 ppm 23 ele	0.014 (0.001)
0.01 ppm 23 ele	0.019 (0.010)
0.1 ppm 23 ele	0.106 (0.100)
1 ppm 23 ele	0.969 (1.000)
10 ppm 23 ele	10.003 (10.000)

Intensity [cps] (10^3)



$$f(x) = 19444.3395 * x + -292.9811$$

$$R^2 = 1.0000$$

$$BEC = -0.015 \text{ ppm}$$

$$LoDt = 0.001 \text{ ppm}$$